



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

### Technical Safety Services, LLC

7570 Trade Street  
San Diego, CA 92121

Fulfills the requirements of

### ISO/IEC 17025:2017

In the field of

### TESTING

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 08 April 2028

Certificate Number: AT-3296.01



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

### Technical Safety Services, LLC

7570 Trade Street  
San Diego, CA 92121

Cheryl Andreoli 858-663-1760  
[candreoli@techsafety.com](mailto:candreoli@techsafety.com)

### TESTING

ISO/IEC 17025 Accreditation Granted: **23 March 23, 2026**

Certificate Number: **AT-3296.01** Certificate Expiry Date: **08 April 2028**

#### Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Water Bioburden Testing	USP <1231>	Medical Devices, Pharmaceuticals	Water Filtration System, ISO Class 5 Hoods, Incubators
Biological Indicators	USP <55> USP <1035>	Medical Devices, Pharmaceuticals	ISO Class 5 Hoods, Incubators
Incubation and Enumeration	USP <797>, USP <1116>, USP <61>	Environmental Monitoring	ISO Class 5 Hoods, Incubators
Organism Identification (Genetic and Gram Stain)	USP <1113>, USP <797>, USP <1116>	Medical Devices, Pharmaceuticals	MALDI-TOF, ISO Class 5 Hoods, Incubators, Microscopes
Bacterial Endotoxins	USP <85 v41>	Medical Devices, Pharmaceuticals	Microplate Reader

#### Chemical

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Water Purity Analysis <ul style="list-style-type: none"> <li>• TOC</li> <li>• Conductivity</li> </ul>	USP <643>, USP <645>	Medical Devices, Pharmaceuticals	TOC/Conductivity Analyzer

Jason Stine, Vice President

This Scope of Accreditation, version 003, was last updated on: 23 March 2026 and is valid only when accompanied by the Certificate.

